Neuroakashic Light Device Characterization Before Evaluation of Human Experience

Ana Silvia Lara
Independent researcher | Ana Silvia Lara Publishing LLC
Indian Wells, California, United States
Correspondence: info@anasilvialara.com

Methodological development paper
Working paper | Released 4 November 2026 | Not peer reviewed

Abstract

Background. Neuroakashic Light combines LED illumination, geometric codes and a capsule described as containing black tourmaline, blue obsidian and white quartz.

Objective. To identify what must be documented about the actual device and its exposure before interpreting reported effects.

Approach. Comparison of the technical sheet, benefits development document and founding dossier, followed by a staged characterization proposal.

Evidence status. Historical descriptions include indicative values and asserted mechanisms, but no calibration report or device-specific efficacy dataset was supplied in the reviewed sources.

Conclusion. A traceable device specification and component comparison are necessary foundations for subsequent experience research.

Keywords: LED illumination; optical exposure; device characterization; reproducibility; wellness technology

Introduction

The author's technical material describes a light-based experience incorporating geometric codes and a mineral capsule [2, 3]. The symbolic roles assigned to the three materials include protection, organization and expansion. Those roles belong to the originating framework and are preserved as such. This paper asks what the actual device emits, under what conditions and with what variation. It does not convert symbolic descriptions into verified optical or biological properties.

Documentary findings and unresolved specifications

The benefits document lists illustrative white, red and near-infrared ranges, irradiance, illuminance and flicker values [3]. Those entries are explicitly treated here as indicative development values. No supplied measurement establishes that the current model has a near-infrared channel, a specified spectral distribution, a low-flicker result or compliance with a photobiological safety standard. The technical sheet's frequency and color associations require an identified optical measurement before use as specifications [2]. Numerical values and mechanism claims remain in their original sources; this manuscript does not silently certify them.

Proposed physical characterization

Identify the base lamp, model, supplier, unit identifier, power source, driver, optical components, capsule configuration and firmware if applicable. Document each code's material, dimensions, position and pattern. Record the emission spectrum, optical output at declared use positions, distance, angle, session duration, temporal variation and temperature. Instruments, calibration, uncertainty and ambient conditions must accompany the records. Mineral identity and optical transmission require their own methods; a retail material name is not an independent composition analysis. Lux and irradiance represent different quantities and cannot be substituted without a justified spectral calculation.

Context from published light research

Brainard and colleagues investigated wavelength-dependent melatonin suppression in humans [4]. Chang and colleagues examined evening light-emitting eReader use and found adverse sleep-related effects under their experimental conditions [5]. These studies show why timing, spectrum and exposure matter; neither tested Neuroakashic Light. They do not establish that its exposure improves sleep or worsens it. A photobiomodulation explanation would similarly require matching the actual device parameters, tissue exposure and research question to relevant evidence rather than borrowing effects from another lamp.

Component comparisons and later experience research

An initial bench comparison could hold lamp settings and geometry constant while varying the capsule or pattern. Repeated measurements across units and sessions would assess reproducibility before a human benefit question is addressed. A later experience study must specify whether it evaluates the combined room experience or the separate contribution of light. Audio, facilitator attention, rest, expectations and environment are potential contributors. A claim about the capsule requires an appropriate capsule comparison. No human dose or treatment regimen is prescribed by this paper; exposure and applicable safety assessment must be established first.

Discussion and conclusion

Characterization is valuable even if a proposed mechanism is not retained. It permits comparison between device versions and makes future research interpretable. The reviewed sources do not establish coherent wavefront generation by ordinary LED illumination, cellular regeneration, neuronal reordering or a therapeutic effect. Those historical statements require distinct evidence and should remain labeled as claims or hypotheses until tested. The immediate deliverable is a verified specification with uncertainty and traceable measurement records.

Conceptual sources and lineage

De Akashico a Neuroakashico places light, codes, intention and observation within the author's conceptual development [6]. The author identifies Manifestación acelerada as a further source of that experiential vocabulary [7]. These citations document provenance and intended meaning; device output, safety and any biological effects still require the measurements described in this paper.

Declarations

Ana Silvia Lara develops the concepts, methods and products discussed and is associated with the publisher. No independent editorial assessment, funding award or institutional endorsement is claimed. No participant dataset was analyzed for this draft. Proposed studies have not been conducted through this work. The author is responsible for the final work. © 2026 Ana Silvia Lara. All rights reserved in protectable original human-authored contributions.

References

[1] Lara AS. Neuroconsciousness Journal Editorial March 2027. English design and manuscript review proof. Version 2; November 2026. Unpublished development dossier; relevant article identified in text.

[2] Lara AS. Ficha tecnica de Neuroakashic Light. 8-page technical and conceptual development document; copyright 2025. Internal source; claimed mechanisms and numerical values are not independently verified measurements.

[3] Lara AS. Beneficios vibracionales y cientificos de Neuroakashic Light to review. 5-page development document; copyright 2025. Internal source; includes explicitly indicative parameters.

[4] Brainard GC, Hanifin JP, Greeson JM, Byrne B, Glickman G, Gerner E, Rollag MD. Action spectrum for melatonin regulation in humans: Evidence for a novel circadian photoreceptor. Journal of Neuroscience. 2001;21(16):6405-6412. https://doi.org/10.1523/JNEUROSCI.21-16-06405.2001

[5] Chang AM, Aeschbach D, Duffy JF, Czeisler CA. Evening use of light-emitting eReaders negatively affects sleep, circadian timing, and next-morning alertness. Proceedings of the National Academy of Sciences. 2015;112(4):1232-1237. https://doi.org/10.1073/pnas.1418490112

[6] Lara AS. De Akashico a Neuroakashico: El nacimiento de la Luz Ilimitada. Ana Silvia Lara Publishing LLC; 2026. Author-supplied paperback manuscript, pp. 190-195.

[7] Lara AS. Manifestación acelerada. n.d. Book identified by the author as a conceptual source; edition and publication year not supplied.